

---

# Colorimetric Determination Of An Equilibrium Constant In Aqueous Solution

---

*Colorimetric  
Determination  
Of An  
Equilibrium  
Constant In  
Aqueous  
Solution*

Downloaded from  
[api.lacavedespapilles.com](http://api.lacavedespapilles.com)  
by Leticia & Dyer

---

## INTRODUCTION TO COLORIMETRIC EQUILIBRIUM ANALYSIS

---

Understanding chemical equilibrium is a cornerstone of physical chemistry, and the ability to quantify equilibrium constants in solution is essential for fields ranging from biochemistry to environmental science.

Among the most accessible and elegant methods for determining an equilibrium constant in aqueous solution is colorimetry. This approach leverages the simple relationship between the concentration of a colored species and the intensity of light it absorbs. By measuring how much light a solution absorbs at a specific wavelength, we can deduce concentrations and, through careful

stoichiometric reasoning, calculate the equilibrium constant for a reversible reaction. This article provides a comprehensive walkthrough of the theory, experimental procedure, data analysis, and practical applications of the colorimetric determination of an equilibrium constant in aqueous solution, using the classic reaction between iron(III) ions and thiocyanate ions as a guiding example.

---


## THE THEORETICAL FOUNDATION: CHEMICAL EQUILIBRIUM AND COLORIMETRY

---

# Chemical

# Equilibrium and the Equilibrium Constant

In a reversible chemical reaction, equilibrium is reached when the rates of the forward and reverse reactions become equal, and the concentrations of reactants and products remain constant over time. For a general reaction in aqueous solution:



the equilibrium constant  $K_c$  (expressed in terms of molar

concentrations) is defined as:

$$K_c = \frac{[C]^c [D]^d}{[A]^a [B]^b}$$

The value of  $K_c$  is constant at a given temperature and provides quantitative insight into the position of equilibrium. A large  $K_c$  indicates that products are favored at equilibrium, while a small  $K_c$  indicates reactants are favored. In the context of colorimetric determination, we must be able to measure at least one species' concentration directly at equilibrium, and from that, calculate all others using initial concentrations and the reaction stoichiometry.

## Principles of Colorimetry and the Beer-Lambert Law

Colorimetry is based on the principle that many substances—particularly transition metal complexes and organic dyes—absorb visible light. The amount of light absorbed is proportional to the concentration of the absorbing species and the path length of the light through the solution. This

relationship is formalized by the Beer-Lambert Law:

$$A = \epsilon b c$$

Where:

- **A** = absorbance (dimensionless)
- **$\epsilon$**  = molar absorptivity ( $\text{L mol}^{-1} \text{cm}^{-1}$ ), a constant for a given substance at a specific wavelength
- **b** = path length of the cuvette (usually 1 cm)
- **c** = concentration of the absorbing species ( $\text{mol L}^{-1}$ )

For the colorimetric determination of an equilibrium constant, we first need to establish a calibration curve: a plot of absorbance versus known concentrations of the colored species.


Then, at equilibrium, we measure the absorbance of the solution and use the calibration curve to find the equilibrium concentration of that species. This concentration is the key that unlocks the entire equilibrium constant calculation.

---

### THE CLASSIC MODEL SYSTEM: **$\text{Fe}^{3+}$ AND $\text{SCN}^-$**

---

The most widely taught laboratory experiment for this topic uses the reaction between iron(III) ions and thiocyanate ions to form the blood-red iron(III) thiocyanate complex:



This reaction is ideal for colorimetric analysis because:

- The product,

$\text{FeSCN}^{2+}$ , is intensely colored (deep red-orange), while the reactants  $\text{Fe}^{3+}$  (pale yellow) and  $\text{SCN}^-$  (colorless) absorb very little visible light at the wavelength of maximum absorption for the complex (around 447–470 nm).

- The reaction is reversible and reaches equilibrium quickly at room temperature.
- The stoichiometry is 1:1, simplifying calculations.

Other possible systems include the copper(II)-ammonia complex or the iron(III)-salicylate complex, but the  $\text{Fe}^{3+}/\text{SCN}^-$  system

remains the standard for introductory and advanced courses alike.

## **EXPERIMENTAL METHODOLOGY FOR EQUILIBRIUM CONSTANT DETERMINATION**

# Materials and Instrumentation

To perform a colorimetric equilibrium constant determination, you will need:

- A spectrophotometer or colorimeter capable of measuring absorbance in

the visible range  
(400–600 nm)

- Standard volumetric glassware (volumetric flasks, pipettes, burettes)
- Solutions of known concentrations: typically 0.002 M  $\text{Fe}(\text{NO}_3)_3$  in 1 M  $\text{HNO}_3$  (the nitric acid suppresses hydrolysis of  $\text{Fe}^{3+}$ ) and 0.002 M KSCN (or NaSCN)
- Deionized water for dilutions
- Cuvettes (usually 1 cm path length)

## Step 1: Preparati

# on of Standard Solutions for the Calibration Curve

A calibration curve is constructed using solutions that contain only the product,  $\text{FeSCN}^{2+}$ , at known concentrations.

However, because the equilibrium constant is not extremely large, it is difficult to prepare pure product solutions. Instead, we use the "reactant excess" method: prepare solutions where  $\text{SCN}^-$  is the limiting reagent and  $\text{Fe}^{3+}$  is in large excess. Under these

conditions, the equilibrium shifts far to the right, and essentially all  $\text{SCN}^-$  is converted to  $\text{FeSCN}^{2+}$ . Thus, the concentration of  $\text{FeSCN}^{2+}$  can be assumed equal to the initial concentration of  $\text{SCN}^-$ .

For example, prepare six standard solutions in 25 mL volumetric flasks: add 5.00 mL of 0.20 M  $\text{Fe}(\text{NO}_3)_3$  (excess) and varying volumes (e.g., 0.50, 1.00, 1.50, 2.00, 2.50, 3.00 mL) of 0.002 M KSCN, then dilute to the mark with 1 M  $\text{HNO}_3$ . The final concentration of  $\text{FeSCN}^{2+}$  in each flask equals (volume KSCN  $\times$  0.002 M) / total volume.

## Step 2: Measuring Absorbance and Building the Calibration Curve

Set the spectrophotometer to the wavelength of maximum absorbance for  $\text{FeSCN}^{2+}$  (typically 447 nm). Zero the instrument using a blank solution (1 M  $\text{HNO}_3$ ). Measure the absorbance of each

standard solution. Plot absorbance (y-axis) versus concentration of  $\text{FeSCN}^{2+}$  (x-axis). The plot should be linear through the origin, confirming the Beer-Lambert Law. Determine the slope of the best-fit line, which equals  $\epsilon \times b$  (with  $b = 1 \text{ cm}$ , slope =  $\epsilon$ ).

## Step 3: Preparing Equilibrium Mixtures

Now prepare a series of equilibrium mixtures where  $\text{Fe}^{3+}$  and  $\text{SCN}^-$  are both at low, comparable initial concentrations. For example, in separate 25 mL flasks, mix

volumes of 0.002 M  $\text{Fe}(\text{NO}_3)_3$  and 0.002 M KSCN such that the initial concentrations are roughly in the range of  $1\text{--}4 \times 10^{-4} \text{ M}$ . Do not add excess  $\text{Fe}^{3+}$ . Dilute each to the mark with 1 M  $\text{HNO}_3$ . Allow the solutions to stand for a few minutes to ensure equilibrium is reached.

## Step 4: Measuring Equilibrium Absorbance

Measure the absorbance of each

equilibrium mixture at the same wavelength used for the calibration curve. Record the absorbance values precisely.

### **DATA ANALYSIS: CALCULATING THE EQUILIBRIUM CONSTANT**

Determini  
ng the  
Equilibriu  
m  
Concentr  
ation of  
the

## Colored Product

Using the calibration curve (or the Beer-Lambert equation), convert the measured absorbance of each equilibrium mixture to the concentration of  $\text{FeSCN}^{2+}$  at equilibrium:

$$[\text{FeSCN}^{2+}]_{\text{eq}} = \frac{\text{Absorbance}}{\text{slope}}$$

Calculatin  
g  
Equilibriu  
m  
Concentr

# ations of g the Reactants Equilibriu m

From the stoichiometry of the reaction (1:1:1), the amount of  $\text{Fe}^{3+}$  and  $\text{SCN}^-$  consumed equals the amount of  $\text{FeSCN}^{2+}$  formed. Therefore:

$$\begin{aligned} & [\text{Fe}^{3+}]_{\text{eq}} = [\text{Fe}^{3+}]_{\text{initial}} - [\text{FeSCN}^{2+}]_{\text{eq}} \\ & [\text{SCN}^-]_{\text{eq}} = [\text{SCN}^-]_{\text{initial}} - [\text{FeSCN}^{2+}]_{\text{eq}} \end{aligned}$$

Where the initial concentrations are calculated from the volumes and stock concentrations used in each equilibrium mixture, accounting for dilution in the 25 mL flask.

## Computin

## Constant

For each equilibrium mixture, calculate  $K_c$  using:

$$K_c = \frac{[\text{FeSCN}^{2+}]_{\text{eq}}}{([\text{Fe}^{3+}]_{\text{eq}} \times [\text{SCN}^-]_{\text{eq}})}$$

Because the reaction is 1:1, the units of  $K_c$  will be  $\text{L mol}^{-1}$ . Compute the average  $K_c$  value from all mixtures. At room temperature (25 °C) and in 1 M  $\text{HNO}_3$ , the expected  $K_c$  for the  $\text{FeSCN}^{2+}$  complex is typically around  $200\text{--}300 \text{ L mol}^{-1}$ , though it varies with ionic strength and acid concentration.

# Error Analysis and Refinements

Replicate measurements and multiple equilibrium mixtures help reduce random error. Systematic errors may arise from:

- Incomplete conversion in the standard solutions (if  $\text{Fe}^{3+}$  excess is insufficient). Using at least 50-fold excess of  $\text{Fe}^{3+}$  ensures >99% conversion of  $\text{SCN}^-$ .
- Absorbance

contributions from  $\text{Fe}^{3+}$  itself. This is minimized by using a blank that contains the same concentration of  $\text{Fe}^{3+}$  as in the samples, or by subtracting a small baseline absorbance.

- Variations in temperature and ionic strength. Performing all measurements at the same temperature and maintaining high and constant ionic strength (via 1 M  $\text{HNO}_3$ ) improves precision.

---

## PRACTICAL CONSIDERATIONS AND TROUBLESHOOTING

## Choosing the Right Wavelength

If using a spectrophotometer, first scan the absorbance spectrum of a representative  $\text{FeSCN}^{2+}$  solution to identify  $\lambda_{\text{max}}$ . For a colorimeter, select a filter that offers maximum absorbance. Avoid wavelengths where the blank or reactants absorb significantly.

## Handling

## Hydrolysis and Side Reactions

$\text{Fe}^{3+}$  ions hydrolyze in water to form  $\text{FeOH}^{2+}$  and other species, which can interfere. The use of 1 M  $\text{HNO}_3$  maintains the solution strongly acidic, suppressing hydrolysis and ensuring that  $\text{Fe}^{3+}$  remains as the hexaaqua ion, which does not absorb significantly at the analysis wavelength.

## Instrument Care

Always use clean, scratch-free cuvettes. Handle cuvettes by the

frosted sides to avoid fingerprints on the optical windows. Rinse cuvettes with the solution to be measured before filling, to eliminate dilution errors.

### VARIATIONS AND EXTENSIONS OF THE METHOD

The same colorimetric approach can be applied to other equilibrium systems. For example, the formation of the copper-ammonia complex  $\text{Cu}(\text{NH}_3)_4^{2+}$  (deep blue) can be used to determine the stepwise formation constants. In that case, the equilibrium is more complex, and multiple species may absorb, requiring careful selection of wavelength or use of a diode-array spectrophotometer.

For reactions with 1:2 or other stoichiometries, the stoichiometric relationships change accordingly. The fundamental principle remains: measure the concentration of one species colorimetrically at equilibrium, then use initial concentrations and the balanced equation to calculate all equilibrium concentrations and thus  $K_c$ .

Advanced variants include the use of **spectrophotometric titration**, where absorbance is recorded as a function of added titrant, and the equilibrium constant is extracted from the inflection point or by nonlinear regression of the entire titration curve. This is particularly useful

when the absorbing product in a weak acid-  
species is a reactant or base